




1695 West MacArthur Boulevard Costa Mesa, California 92626  
Tel: (714) 32700600 Fax: (714) 327-0610  
eMail: lambda@lambda.cc http://www.lambda.cc

## Certificate of Conformance

11/18/2020

Order Number : C61267/2  
Customer PO : S475660  
Customer Name : CALTECH  
Part Number : 44791  
Description : HHR-PCCM-2506U-150-1064-10  
Customer Part # :  
Quantity : 2  
Lot Number : S1:C11111220 ( HHR )  
S2:C31111120 ( AR )  
Specification : Material: UV Fused Silica  
Diameter: 25.4mm Tol: +0.00/-0.20mm  
Thickness: TE=6.35mm Tol: +/-0.25mm  
R1: 150mm +/- 0.5%  
Surface Figure: lambda/10 @ 633nm  
Surface Quality: 10-5 (both sides)  
Centration Error: <= 3 arcmin  
Bevel: 0.3mm @ 45 Deg.  
CA: >=85% of center diameter  
  
COATING:  
S1 (CC): HHR R>99.5% @1064nm, 10 deg p-pol.  
S2 (PO): AR R<0.20% @ 1064nm, 10 deg p-pol

The above items have passed quality control inspection and are manufactured to these specifications, unless otherwise indicated below

  
QC Manager, Joe Defilippo

11/18/20  
Date

# Lambda Research Optics Inc.


Date: 11/17/20

Lot#: C1-1-111220

Customer: CALTECH

Coating Type: HHR R>99.5% @1064nm, 10°, P-Pol

Job#: C6126712




1111220P.SP - 11/17/20

Initial: DP

ccdept  
Wednesday, November 11, 2020 12:04 PM

Analyst  
Date



Lot : C31111120  
Coating : AR R < 0.20% @ 1064 nm 10° p-pol  
Customer : CALTECH

C31111120